

O. Laurent¹, C. Philippon¹, C. Yver Kwok¹, L. Rivier¹, M. Ramonet¹
¹Laboratoire des Sciences du Climat et de l'Environnement (LSCE)

N2O Analyzers tested

ICOS compliant analyzer for N2O measurement :



Picarro G5310



LGR CO/N2O EP



ECOTECH Spectronus FTIR

Technology	Mid IR CRDS	Mid IR OA-ICOS	FTIR
Species	N ₂ O/CO/H ₂ O	N ₂ O/CO/H ₂ O	N ₂ O/CO/CO ₂ /CH ₄ /δ ¹³ C-CO ₂ /δ ¹⁸ O-CO ₂ /H ₂ O
Maintenance	coolant refill / 2 months	No maintenance	N ₂ tank every 1 or 2 month, Mg perchlo
Dryer	Required	Required	Integrated (Nafion + Mg(ClO ₄) ₂)
Short Term Working standard (STWS)	At least 1 / day	At least 2/day	

Uncertainty Source

The N2O measurement **uncertainty** (bias and precision) **sources** :

- Temperature sensitivity
- Atmospheric pressure sensitivity
- H₂O measurement offset variability/drift
- Intrinsic variability (electronics, optics...)
- H₂O correction precision/bias
- H₂O correction drift
- H₂O sensitivity (slope) variability
- Linearity (calibration fit residuals)
- Spectroscopic cross sensitivity with other species
- Uncertainty of assigned value of calibration scale
- Artifacts from Sampling system upstream analyzer

Improvement actions

- Can be corrected with a **Short Term Working Standard (STWS)** measured regularly
- Dry ambient air sample
- Optimize the **calibration fit order**
- Mlab assessment
- Optimize procedure (CAL FCL)
- Regular assessment (shelter test...)


Once the improvement actions performed, the remaining uncertainties must be assessed

WMO Network compatibility goal : 0.1 ppb N₂O


Precision and Bias

Long Term Repeatability (LTR) = **Target injection SD**

At ICOS Metrological Lab




In the Field (several months)




Water vapor

Water vapor correction error :



Picarro G5301 analyzer




Any H2O variability (e.g. temperature variability around Nafion dryer) in the bend area induce a N2O error variability.

Note: Drying technique such as Nafion induce a small bias (up to -0,3 ppb depending for Nafion PD model). However passing Calibration gas through Dryer should correct most of the bias.

	NO dryer Factory	NO dryer ATC	Cryo	Nafion Only for amb. air	Nafion For Cal. & amb. air
Picarro	0.6	0.3	0	0.2	0.05
LGR	0.6	0.15	0	0.15	0.05

Linearity



Applied Calibration fit order: **1**



CO₂ cross sensitivity

Picarro

LGR



	Picarro	LGR	FTIR
Linearity error (ppb) Typical/Max	0.03/0.1	0.03/0.1	0.05/0.15


Picarro and LGR show a different Linearity (cross sensitivity with another air species?). An appropriated calibration fit order can induce a significant error.

Validation in ambient air


Hourly average data STWS corrected.

N2O Reference: Picarro with cryo and 5h STWS

Cryo dryer (-60°C)



Nafion dryer



Uncertainty Budget

	Picarro	LGR	FTIR
H2O correction error with Nafion (ppb) Typical/Max	0.02/0.05	0.02/0.05	0
Remaining Nafion artifacts (ppb) Typical/Max	0.01/0.03	0.01/0.03	0.01/0.03
Linearity error (ppb) Typical/Max	0.03/0.1	0.03/0.1	0.05/0.15
Precision with STWS correction (ppb) Typical/Max	0.02/0.07	0.07/0.15	0.05/0.15
Overall Uncertainty (ppb) Typical/Max (Quadrature sum)	0.04/0.13	0.08/0.19	0.07/0.21

Minute average data